

PROSTHETIST: _____ PO NUMBER: _____ Date: _____

Trans-Tibial

1. A modified FLO-TECH-TOR™ made from 3/16" firm Pelite™ and Velcro®
2. A standard FLO-TECH® neoprene thigh band
3. An auxiliary neoprene band applied at the knee
4. A standard 9" FLO-TECH® neoprene socket band
5. A removable, anatomically correct support strut, attachment band for strut
6. A neoprene distal socket retainer (cap)
7. Standard 1" and 2" distal end pads.

Regular:

10" length (6¾" to 10") **One Size** **L** _____ **R** _____
 Distal Range: 11" (28cm) to 23" (58.5cm)

Extra Large

10" length (6¾" to 10") **One Size** **L** _____ **R** _____
 Distal Range: 20" (51cm) to 28" (71cm)

Trans-Femoral

1. A modified FLO-TECH-TOR™ made from 3/16" firm Pelite™ and Velcro®
2. A standard FLO-TECH® neoprene 2-piece pelvic band
3. A 9" lateral socket extension band (used to extend the lateral trochanter area of the socket to interface with the trochanter section of the pelvic band.
 - a. A 2" by 5" double-sided Velcro® pressed between a standard neoprene trochanter section of the pelvic belt and the lateral socket extension band.
 - b. The lateral socket extension band may be used either horizontally or vertically depending on the size of the patient.
 - c. It is attached to the SOFTIE™-TF with 2 patches of Velcro® at the proximal lateral area
4. A standard 9" FLO-TECH-TOR™-TF neoprene socket band
5. A neoprene distal socket retainer (cap)
6. Standard 1" and 2" distal end pads.

Regular:

13" length (9¾" to 13") **One Size** **Universal Side** **QTY** _____
 Distal Range: 13" (33cm) to 28" (71cm)

Extra Large

13" length (9¾" to 13") **One Size** **Universal Side** **QTY** _____
 Distal Range: 23" (58.5cm) to 38" (96.5cm)

COMFORT OPTIONS & COMPONENT ITEMS:

<u>QTY</u>	<u>DESCRIPTION</u>	<u>QTY</u>	<u>DESCRIPTION</u>
_____	Trans-Tibial Waist Belt/Suspension Strap	_____	Each Amputation Socks Size: _____
_____	Fork/Suspension Strap	_____	Distal End Pad – per pkg. Size: _____
_____	Neoplush Pelvic Band (Amputated side)	_____	Distal End Pad Reticulated Size: _____
_____	Neoplush Pelvic Band (Contralateral side)		
_____	Neoplush Socket Band		



FLO-TECH® O & P Systems, Inc.
 PO Box 462 - 7325 Halseyville Road
 Trumansburg, New York 14886
 www.1800FLO-TECH.com
 All materials © FLO-TECH® Modified March 2017

Phone: 1-800-FLO-TECH (356-8324)
Phone: (607) 387-3070
Fax: (607) 387-3176
E-mail: info@1800FLO-TECH.com